

Flow retrofit service product

Product type plate instruction



Correct information allows effective retrofit procedures

Introduction

This document highlights the key considerations related to updating the product name plate as part of the retrofit procedure.

A product retrofitted with a new electronics transmitter will require a new name plate containing information about the existing sensor and information about the new electronics transmitter.

Why is it important to update the data label?

A data label is required to provide permanent product identification for a number of key reasons:

Legal

- Country of origin and manufacturers details are required under international law, including CE regulations

Safety information for plant and service staff

- Protection class information
- Power supply rating
- Operating temperature and process information
- Product identification

Preparation of a new data label

ABB can prepare the labels and supply either with the kit of parts or separately.

Please provide your ABB representative with the serial number or a photograph of your existing product label – see example label above.

Flow services



ABB Limited
Measurement & Analytics
Oldends Lane, Stonehouse
Gloucestershire GL10 3TA, UK
Tel: +44 (0)1453 826661
Fax: +44 (0)1453 829671

abb.com/measurement

We reserve the right to make technical changes or modify the contents of this document without prior notice. With regard to purchase orders, the agreed particulars shall prevail. ABB does not accept any responsibility whatsoever for potential errors or possible lack of information in this document.

We reserve all rights in this document and in the subject matter and illustrations contained therein. Any reproduction, disclosure to third parties or utilization of its contents – in whole or in parts – is forbidden without prior written consent of ABB.
©ABB 2018 3KXS240011R2001